

Routine Orbit SUMMARY

Indications	Trauma, Pain, Swelling																									
Diagnostic Task	Detect fractures, edema, masses, or infection of the eye																									
Scan mode	Helical																									
Position/Landmark	Head -Supine																									
Helical Set	<table style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 5%;"></th> <th style="width: 30%;">body part</th> <th style="width: 20%;">thickness spacing</th> <th style="width: 25%;">algorithm</th> <th style="width: 20%;">recon destination</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>orbit bones</td> <td>0.6mmx 0.6mm</td> <td>bone</td> <td>pacs</td> </tr> <tr> <td>2</td> <td>orbit soft tissue</td> <td>2mmx 2mm</td> <td>standard</td> <td>pacs</td> </tr> <tr> <td>3</td> <td>sag orbit soft tissue</td> <td>2mmx2mm</td> <td>standard</td> <td>pacs</td> </tr> <tr> <td>4</td> <td>coronal orbit soft tissue</td> <td>2mmx2mm</td> <td>standard</td> <td>pacs</td> </tr> </tbody> </table>		body part	thickness spacing	algorithm	recon destination	1	orbit bones	0.6mmx 0.6mm	bone	pacs	2	orbit soft tissue	2mmx 2mm	standard	pacs	3	sag orbit soft tissue	2mmx2mm	standard	pacs	4	coronal orbit soft tissue	2mmx2mm	standard	pacs
		body part	thickness spacing	algorithm	recon destination																					
	1	orbit bones	0.6mmx 0.6mm	bone	pacs																					
	2	orbit soft tissue	2mmx 2mm	standard	pacs																					
	3	sag orbit soft tissue	2mmx2mm	standard	pacs																					
4	coronal orbit soft tissue	2mmx2mm	standard	pacs																						
Scan Start/end location	1cm superior to frontal sinus																									
	through maxilla																									
DFOV	25cm																									
angle	none																									
IV contrast volume/type	80ml under 250lbs 100ml over 250lbs isovue 370 2cc/sec if needed																									
Scan delay	60 seconds																									

Mark rt side of face with BB.